

**COVID-19 Antigen Test**

**Clinical Sensitivity and Specificity Study Report**

Drafted by: Nicolas Ma

Approved by: Jerry Zheng

Final report date: 20201216



**Management of the study: Artron Laboratories Inc.**

R& D Department  
Quality Control Department



## Table of Contents

Title Page

Table of Contents

Study Summary

1. Purpose
2. Reference and Compliance
3. Materials
4. Experiment Design
5. Evaluation Criteria
6. Result
7. Conclusion
8. Report



## Study Summary

The purpose of this study was to obtain accurate information regarding the clinical performance Sensitivity and Specificity of Artron COVID-19 Antigen Test. The clinical evaluation was carried out for the clinical performance of Artron COVID-19 Antigen Test.

### 1. Purpose

To validate the clinical performance sensitivity and specificity of Artron COVID-19 Antigen Test.

### 2. Reference and Compliance

FDA guidance for In vitro diagnostic medical device  
WHO Instructions and requirements for Emergency Use Listing (EUL)  
The present study conformed to all applicable laws and regulations.

### 3. Materials

- COVID-19 positive specimens were confirmed by RT-PCR. And non-COVID-19 Nasal/Nasopharyngeal Swabs from non-febrile and non-respiratory patients and confirmed non-COVID-19 by RT-PCR.
- Artron COVID-19 Antigen Test, Lot: COV19-Ag-S-1

### 4. Study Design:

4.1 The clinical performance was evaluated in RT-PCR confirmed COVID-19 infected Nasopharyngeal swab specimens and non-COVID-19 infected Nasopharyngeal Swabs specimens from subjects in the chosen hospitals and clinical laboratories. All the samples should be tested with Artron COVID-19 AntigenTest. 4.2 Test conditions:

- All tests were performed by the clinical technicians in the clinical laboratory according to the manufacturer's instructions using the confirmed samples.
- Visual interpretations of the results of COVID-19 Antigen Test were made independently by the clinical technician.
- The testing center was responsible to summarize the result.

**5. Evaluation Criteria**

Positive (POS): Both C and T lines appear regardless of color intensity. Negative

(NEG): Only C line appears.

Indeterminable (IND): C line appears; T line is so weak that it cannot be determined as positive.

Invalid result (/): Neither C line nor T line appears.

**6. Clinical Results**

**6.1 Clinical Site One**

Experiments have been done at Santa Ana Hospital, Manila, Philippines

The clinical evaluation was carried out for the clinical performance of Artron COVID-19 Antigen Test. Total 38 nasopharyngeal swabs from COVID-19 infected patients and 133 non-COVID-19 infected patients were tested.

Table 1 Artron COVID-19 Antigen Test results

|          |          | Results of COVID 19 Ag test |          | Subtotal |
|----------|----------|-----------------------------|----------|----------|
|          |          | Positive                    | Negative |          |
| RT-PCR   | Positive | 37                          | 1        | 38       |
|          | Negative | 0                           | 133      | 133      |
| Subtotal |          | 37                          | 134      | 171      |

Artron COVID 19 Ag Test kit could detect the Nasopharyngeal swabs with high clinical performance sensitivity and specificity listed below:

Sensitivity:  $37/38 = 97.37\%$  (95% CI: 86.19% – 99.93%)

Specificity:  $133/133 = 100.0\%$  (95% CI: 97.26% – 100%)

Total Agreement:  $(37+133)/171 = 99.42\%$

Total 38 Nasopharyngeal swabs from SARS-Cov-2 infected patients, and 133 non-SARS-Cov-2 infected patients were tested. Among all of those chosen samples, Artron COVID-19 Ag test kit identified out total 37 SARS-Cov-2 positive from 38 SARS-Cov-2 infected patients' samples; the clinical sensitivity was 97.37%, no false positive from total 133 non-SARS-Cov-2 infected cases, which demonstrated the specificity was 100%, total agreement is 99.42%



## 6.2 Clinical Site Two

Experiments have been done in local private hospitals/labs in Ahmedabad, Gujarat, India.

The clinical evaluation was carried out for the clinical performance of Artron COVID-19 Antigen Test. Total 56 Nasopharyngeal swabs from COVID-19 infected patients and 68 non-COVID-19 infected patients were tested.

Table 2 Artron COVID-19 Antigen Test results

|          |          | Results of COVID 19 Ag test |          | Subtotal |
|----------|----------|-----------------------------|----------|----------|
|          |          | Positive                    | Negative |          |
| RT-PCR   | Positive | 48                          | 8        | 56       |
|          | Negative | 0                           | 68       | 68       |
| Subtotal |          | 48                          | 76       | 124      |

Artron COVID 19 Ag Test kit could detect the Nasopharyngeal swabs with high clinical performance sensitivity and specificity listed below:

Sensitivity:  $48/56 = 85.71\%$  (95% CI: 73.78% – 93.62%)

Specificity:  $68/68 = 100.0\%$  (95% CI: 94.72% – 100%)

Total Agreement:  $(48+68)/124 = 93.55\%$

Total 56 Nasopharyngeal swabs from SARS-Cov-2 infected patients, and 68 non-SARS-Cov-2 infected patients were tested. Among all of those chosen samples, Artron COVID-19 Ag test kit identified out total 48 SARS-Cov-2 positive from 56 SARS-Cov-2 infected patients' samples; the clinical sensitivity was 85.71%, no false positive from total 68 non-SARS-Cov-2 infected cases, which demonstrated the specificity was 100%, total agreement is 93.55%

## 6.3 Clinical Site Three

Experiments have been done in Hemolab Clinic, Bucharest, Romania.

The clinical evaluation was carried out for the clinical performance of Artron COVID-19 Antigen Test. This was a prospective study performed over 2 weeks on 303 total subjects.

Table 3 Artron COVID-19 Antigen Test results

|          |          | Results of COVID 19 Ag test |          | Subtotal |
|----------|----------|-----------------------------|----------|----------|
|          |          | Positive                    | Negative |          |
| RT-PCR   | Positive | 30                          | 2*       | 32       |
|          | Negative | 0                           | 271      | 271      |
| Subtotal |          | 30                          | 273      | 303      |

\*2 inconclusive PCR results counted as presumptive positives

Artron COVID 19 Ag Test kit could detect the Nasopharyngeal swabs with high clinical performance sensitivity and specificity listed below:

Sensitivity:  $30/32 = 93.75\%$  (95% CI: 79.19% – 99.23%)

Specificity:  $271/271 = 100.0\%$  (95% CI: 98.65% – 100%)

Total Agreement:  $(30+271)/303 = 99.34\%$

Table 4 Ct Value for positive RT-PCR Results, Clinical Site Three

|        |         |
|--------|---------|
| Count  | 30      |
| Median | 25      |
| Mode   | 26      |
| Range  | 19 – 31 |

Total 32 Nasopharyngeal swabs from SARS-Cov-2 infected patients, and 271 non-SARS-Cov-2 infected patients were tested. Among all of those chosen samples, Artron COVID-19 Ag test kit identified out total 30 SARS-Cov-2 positive from 32 SARS-Cov-2 infected patients' samples; the clinical sensitivity was 93.75%, no false positive from total 271 non-SARS-Cov-2 infected cases, which demonstrated the specificity was 100%, total agreement is 99.34%

## 7. Conclusion

Clinical performance of the Artron COVID-19 Antigen Test was evaluated at three clinical sites. A total of 598 subjects were tested, 126 of whom were determined to be SARS-Cov-2 infected. The sensitivity and specificity of Artron COVID-19 Antigen Test to the positive Nasopharyngeal swab specimens were calculated based on the detection results of RT-PCR. The Artron COVID-19 Antigen test sensitivity is 91.27% with 95% CI 84.92% – 95.56%, and specificity is 100% with 95% CI 99.22% – 100%. The overall agreement is 98.16 %.

## 8. Report

8.1 Original raw data is archived at Quality Control Department

8.2 The original final report is archived in Quality Control Department.